



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

PYLON ELECTRONICS INC.
147 Colonnade Road
Ottawa, Ontario, Canada K2E 7L9
Adnan Mahdi Phone: 613 226 7920

CALIBRATION

Valid To: September 30, 2026

Certificate Number: 3630.01

In recognition of the successful completion of the A2LA evaluation process (including an assessment of the organization's compliance with A2LA's Calibration Program Requirements), accreditation is granted to this laboratory to perform the following calibrations^{1,7}:

I. Dimensional

Parameter/Equipment	Range	CMC ^{2,4} (\pm)	Comments
Gauge Blocks - Length Only	Up to 1 in (> 1 to 4 in)	8.6 μ in (7.2 + 1.5L) μ in	Gauge block comparator, master blocks
	Up to 20 mm (> 20 to 100 mm)	0.22 μ m (0.19 + 0.0014L) μ m	
Gauge Block Comparator	(0.1 \pm 0.000 05) in	2.2 μ in	Master gauge blocks
Electronic Gaging Unit	\pm 0.01 in	7 μ in	Gauge blocks
Check Master Length ⁹	Up to 24 in	(46 + 2.3L) μ in	Precision height gauge, gauge blocks
Parallelism	Up to 24 in (Height)	11 μ in	Precision height gauge
Parallelism (Micrometer)	Anvil & Spindle	10 μ in	Optical flat

Parameter/Equipment	Range	CMC ^{2, 4} (\pm)	Comments
Hand Tools ^{3, 9} –			
Calipers (OD/ID/Depth)	Up to 40 in (1000 mm)	320 μ in (8.2 μ m)	Gauge blocks
Depth Gages	Up to 24 in (600 mm)	300 μ in (7.7 μ m)	Gauge blocks
Height Gages	Up to 40 in (1000 mm)	(62 + 2.2L) μ in	Gauge blocks
Dial Indicators	Up to 4 in (100 mm) (> 4 to 12) in (300 mm)	(5.7 + 20L) μ in (10 + 44L) μ in	ULM
Thickness Gage	Up to 8 in (200 mm)	(6.5 + 1.5L) μ in	Gauge blocks
Micrometers OD	Up to 24 in (600 mm)	(12 + 4.1L) μ in	Gauge blocks
Micrometers ID	Up to 24 in (600 mm)	(60 + 2.1L) μ in	ULM, gauge blocks
Micrometers Depth	Up to 24 in (600 mm)	(34 + 3.6L) μ in	Gauge blocks
Bore Gauge Tri Micrometer	Up to 4 in (100 mm)	70 μ in	Secondary ring gauges
Flatness ^{3, 9} – Anvils & Spindles	Up to 0.001 in	10 μ in	Optical flat
Plain Ring Gauges ⁹	Up to 8 in (200 mm) (> 8 to 12) in (300 mm)	18 μ in 45 μ in	ULM, gauge blocks
Plain Plug Gauges	Up to 24 in (600 mm)	(7.5 + 3.4L) μ in (0.18 + 0.0034L) μ m	ULM, gauge blocks
Length – 1D ⁹ – Measure	Up to 24 in (600 mm)	(7.5 + 3.4L) μ in (0.18 + 0.0034L) μ m	ULM, precision linear height gauge, gauge blocks
Surface Plate ³	Up to 12x12 ft (204 in diagonal)	(24 + 0.06L) μ in (diagonal length)	Laser interferometer
Local Flatness	Up to 0.001 in	24 μ in	Repeat-o-meter

Parameter/Equipment	Range	CMC ^{2, 4} (\pm)	Comments
Precision Linear Height Gauge ³	Up to 48 in	$(32 + 2.9L) \mu\text{in}$	Gauge blocks

II. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC ^{2, 5} (\pm)	Comments
DC Voltage – Measure & Generate, Fixed Points	100 mV 1 V 10 V 100 V 1000 V	0.70 $\mu\text{V/V}$ 0.46 $\mu\text{V/V}$ 0.40 $\mu\text{V/V}$ 0.46 $\mu\text{V/V}$ 0.70 $\mu\text{V/V}$	732C, 752A, 845AB, 5720A or 5730A, 3458A
DC Voltage – Measure	0 V (0 to 202) mV (0.202 to 2.02) V (2.02 to 20.2) V (20.2 to 202) V (202 to 1050) V (0 to 202) mV (0.202 to 2.02) V (2.02 to 20.2) V (20.2 to 202) V (202 to 1050) V (0 to 120) mV (0.12 to 1.2) V (1.2 to 12) V (12 to 120) V (120 to 1000) V (0 to 10) kV (10 to 140) kV	0.01 μV 0.79 $\mu\text{V/V} + 0.2 \mu\text{V}$ 0.62 $\mu\text{V/V} + 0.3 \mu\text{V}$ 0.63 $\mu\text{V/V} + 0.5 \mu\text{V}$ 0.67 $\mu\text{V/V} + 5 \mu\text{V}$ 0.83 $\mu\text{V/V} + 50 \mu\text{V}$ 7.5 $\mu\text{V/V} + 0.2 \mu\text{V}$ 2.9 $\mu\text{V/V} + 0.3 \mu\text{V}$ 2.9 $\mu\text{V/V} + 0.5 \mu\text{V}$ 4.3 $\mu\text{V/V} + 30 \mu\text{V}$ 4.4 $\mu\text{V/V} + 0.5 \text{mV}$ 10 $\mu\text{V/V} + 0.34 \mu\text{V}$ 9.1 $\mu\text{V/V} + 0.34 \mu\text{V}$ 9.1 $\mu\text{V/V} + 0.57 \mu\text{V}$ 11 $\mu\text{V/V} + 34 \mu\text{V}$ $(11 + (14V/1000)^2) \mu\text{V/V} + 110 \mu\text{V}$ 0.034 % + 0.034 V 0.091 % + 0.8 V	Copper short 732C, 752A, 8588A 8588A 3458A $V = \text{voltage}$ Vitrek 4700 w/ HVL-150 probe

Parameter/Equipment	Range	CMC ^{2, 5} (\pm)	Comments
DC Voltage – Generate	(0 to 220) mV (0.22 to 2.2) V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1100) V	1.1 μ V/V + 0.2 μ V 1.0 μ V/V + 0.59 μ V 0.43 μ V/V + 2.1 μ V 0.40 μ V/V + 4 μ V 1.1 μ V/V + 40 μ V 1.2 μ V/V + 210 μ V	732C, 752A, 5720A or 5730A
	(0 to 220) mV (0.22 to 2.2) V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1100) V	5.4 μ V/V + 0.39 μ V 3.1 μ V/V + 0.62 μ V 2.3 μ V/V + 2.3 μ V 2.3 μ V/V + 3.9 μ V 3.1 μ V/V + 39 μ V 4.7 μ V/V + 390 μ V	5720A or 5730A
DC Voltage ³ – Generate	(0 to 329.9999) mV (0.33 to 3.299 999) V (3.3 to 32.999 99) V (33 to 329.9999) V (330 to 1020.000) V	15 μ V/V + 0.76 μ V 8.4 μ V/V + 1.5 μ V 9.1 μ V/V + 15 μ V 14 μ V/V + 110 μ V 14 μ V/V + 1.1 mV	552X series calibrator
DC Current – Generate & Measure	(0 to 2) pA (0 to 20) pA (0 to 200) pA (0 to 2) nA (0 to 20) nA (0 to 200) nA (0 to 2) μ A (0 to 20) μ A	0.28 % + 0.39 aA 0.11 % + 3.9 aA 0.08 % + 39 aA 0.033 % + 0.39 aA 78 μ A/A + 3.9 fA 18 μ A/A + 39 fA 12 μ A/A + 0.39 pA 10 μ A/A + 3.9 pA	Standard resistors, 5720A or 5730A, Keithley 487
DC Current – Measure	0 A (0 to 20) μ A (0 to 2) mA (0 to 10) mA (0 to 20) mA (0 to 100) mA (0 to 0.5) A (0 to 3) A (0 to 30) A (0 to 100) A (0 to 10) A (10 to 20) A	14 pA 7.5 μ A/A + 21 pA 3.1 μ A/A + 21 pA 3.2 μ A/A + 200 pA 3.2 μ A/A + 2 nA 3.2 μ A/A + 20 nA 3.4 μ A/A + 200 nA 6.0 μ A/A + 2 μ A 12 μ A/A + 5.1 μ A 24 μ A/A + 200 μ A 33 μ A/A + 20 μ A 66 μ A/A + 20 μ A	Open input 742A-10k, 8588A 742A-1k, 8588A 742A-100, 8588A 742A-10, 8588A 742A-1, 8588A VFP-4 (0.1), 8588A 9230-0.04, 8588A CS-100, 8588A Y5020, 8588A

Parameter/Equipment	Range	CMC ^{2, 5} (\pm)	Comments
DC Current – Measure (cont)	(0 to 20.2) μ A (20.2 to 202) μ A (0.202 to 2.02) mA (2.02 to 20.2) mA (20.2 to 202) A (0.202 to 2.02) A (2.02 to 20.2) A (20.2 to 30.2) A	27 μ A/A + 0.4 nA 9.8 μ A/A + 0.4 nA 9.2 μ A/A + 4 nA 14 μ A/A + 40 nA 57 μ A/A + 1 μ A 140 μ A/A + 0.1 mA 240 μ A/A + 0.4 mA 560 μ A/A + 4.4 mA	8588A
DC Current ³ – Measure	(0 to 120) nA (0.12 to 1.2) μ A (1.2 to 12) μ A (12 to 120) μ A (0.12 to 1.2) mA (1.2 to 12) mA (12 to 120) mA (0.12 to 1.05) A (0 to 500) A	34 μ A/A + 45 pA 23 μ A/A + 45 pA 23 μ A/A + 110 pA 23 μ A/A + 0.91 nA 23 μ A/A + 5.7 nA 23 μ A/A + 57 nA 40 μ A/A + 0.57 μ A 120 μ A/A + 11 μ A 0.56 % + 0.04 A	3458A 34401A
DC Current – Generate	(0 to 220) μ A (0.22 to 2.2) mA (2.2 to 22) mA (22 to 220) mA (0.22 to 2.2) A (2.2 to 11) A	33 μ A/A + 5.4 nA 27 μ A/A + 6.2 nA 27 μ A/A + 39 nA 35 μ A/A + 0.62 μ A 54 μ A/A + 12 μ A 260 μ A/A + 370 μ A	5720A or 5730A w/ 5725A
DC Current ³ – Generate	(0 to 329.999) μ A (0.330 to 3.299 99) mA (3.3 to 32.9999) mA (33 to 329.999) mA (0.33 to 1.099 99) A (1.1 to 2.999 99) A (3 to 10.9999) A (11 to 20.5) A	110 μ A/A + 0.015 μ A 76 μ A/A + 0.038 μ A 76 μ A/A + 0.19 μ A 76 μ A/A + 1.9 μ A 150 μ A/A + 30 μ A 290 μ A/A + 30 μ A 380 μ A/A + 380 μ A 760 μ A/A + 570 μ A	552X series calibrator
Clamp-on Meters	(20 to 149.999) A (150 to 549.999) A (550 to 1025) A	0.58 % + 0.17 A 0.58 % + 0.58 A 0.59 % + 0.58 A	552X series w/ 5500A/coil

Parameter/Equipment	Range	CMC ^{2, 5} (\pm)	Comments
Resistance – Measure/Generate	0 Ω (0.1 to 1) Ω 1 Ω (1 to 10) Ω (10 to 100) Ω (0.1 to 1) k Ω (1 to 10) k Ω 10 k Ω (10 to 100) k Ω (0.1 to 1) M Ω (1 to 10) M Ω (10 to 100) M Ω (0 to 2.02) Ω (2.02 to 20.2) Ω (20.2 to 202) Ω 202 Ω to 2.02 k Ω (2.02 to 20.2) k Ω (20.2 to 202) k Ω 202 k Ω to 2.02 M Ω (2.02 to 20.2) M Ω (20.2 to 202) M Ω 202 M Ω to 2.02 G Ω (2.02 to 20.2) G Ω	47 n Ω 0.50 $\mu\Omega/\Omega$ 0.30 $\mu\Omega/\Omega$ 0.35 $\mu\Omega/\Omega$ 0.40 $\mu\Omega/\Omega$ 0.45 $\mu\Omega/\Omega$ 0.35 $\mu\Omega/\Omega$ 0.30 $\mu\Omega/\Omega$ 0.60 $\mu\Omega/\Omega$ 0.90 $\mu\Omega/\Omega$ 2.9 $\mu\Omega/\Omega$ 12 $\mu\Omega/\Omega$ 17 $\mu\Omega/\Omega$ + 4 $\mu\Omega$ 10 $\mu\Omega/\Omega$ + 14 $\mu\Omega$ 9.2 $\mu\Omega/\Omega$ + 50 $\mu\Omega$ 9.1 $\mu\Omega/\Omega$ + 0.5 m Ω 9.2 $\mu\Omega/\Omega$ + 5 m Ω 9.3 $\mu\Omega/\Omega$ + 50 m Ω 11 $\mu\Omega/\Omega$ + 1 Ω 17 $\mu\Omega/\Omega$ + 10 Ω 68 $\mu\Omega/\Omega$ + 1 k Ω 230 $\mu\Omega/\Omega$ + 0.1 M Ω 0.14 % + 10 M Ω	Copper short MI 6242D bridge with standard resistors 8588A with 552X series calibrator or standard resistors
Resistance – Measure Fixed Points	1 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω	6.9 $\mu\Omega/\Omega$ 2.6 $\mu\Omega/\Omega$ 1.4 $\mu\Omega/\Omega$ 1.4 $\mu\Omega/\Omega$ 1.4 $\mu\Omega/\Omega$ 1.9 $\mu\Omega/\Omega$ 2.8 $\mu\Omega/\Omega$ 4.4 $\mu\Omega/\Omega$ 20 $\mu\Omega/\Omega$	742A-1, 8588A 742A-10, 8588A 742A-100, 8588A 742A-1k, 8588A 742A-10k, 8588A Ohm-Labs 105, 8588A Ohm-Labs 106, 8588A 742A-10M, 8588A 9330-100M, 8588A

Parameter/Equipment	Range	CMC ^{2, 5, 6} (\pm)	Comments
Resistance – Measure	(0 to 2.02) Ω (2.02 to 20.2) Ω (20.2 to 202) Ω 202 Ω to 2.02 k Ω (2.02 to 20.2) k Ω (20.2 to 202) k Ω 202 k Ω to 2.02 M Ω (2.02 to 20.2) M Ω (20.2 to 202) M Ω 1 G Ω 10 G Ω (1 to 10) M Ω (10 to 100) M Ω 100 M Ω to 1 G Ω (1 to 10) G Ω (10 to 100) G Ω 100 G Ω to 1 T Ω	5 $\mu\Omega/\Omega$ + 4 $\mu\Omega$ 2.5 $\mu\Omega/\Omega$ + 14 $\mu\Omega$ 1.6 $\mu\Omega/\Omega$ + 49 $\mu\Omega$ 1.1 $\mu\Omega/\Omega$ + 0.47 m Ω 1.1 $\mu\Omega/\Omega$ + 4.8 m Ω 1.6 $\mu\Omega/\Omega$ + 36 m Ω 2.8 $\mu\Omega/\Omega$ + 0.81 Ω 6.5 $\mu\Omega/\Omega$ + 9.1 Ω 15 $\mu\Omega/\Omega$ + 720 Ω 60 $\mu\Omega/\Omega$ 140 $\mu\Omega/\Omega$ 0.029 % 0.040 % 0.057 % 0.080 % 0.11 % 0.23 %	742A-1, 8588A 742A-10, 8588A 742A-100, 8588A 742A-1k, 8588A 742A-10k, 8588A Ohm-Labs 105, 8588A Ohm-Labs 106, 8588A 742A-10M, 8588A 9330-100M, 8588A 9330-100M, 5720A, Keithley 487 Guildline 9520
Resistance ³ – Measure	(0 to 12) Ω (12 to 120) Ω 120 Ω to 1.2 k Ω (1.2 to 12) k Ω (12 to 120) k Ω 120 k Ω to 1.2 M Ω (1.2 to 12) M Ω (12 to 120) M Ω 120 M Ω to 1.2 G Ω	17 $\mu\Omega/\Omega$ + 57 $\mu\Omega$ 14 $\mu\Omega/\Omega$ + 0.57 m Ω 11 $\mu\Omega/\Omega$ + 0.57 m Ω 11 $\mu\Omega/\Omega$ + 5.7 m Ω 11 $\mu\Omega/\Omega$ + 57 m Ω 17 $\mu\Omega/\Omega$ + 2.3 Ω 57 $\mu\Omega/\Omega$ + 110 Ω 570 $\mu\Omega/\Omega$ + 1.1 k Ω 0.57 % + 11 k Ω	3458A

Parameter/Equipment	Range	CMC ^{2, 5, 6} (\pm)	Comments
Resistance – Generate Fixed Points	100 $\mu\Omega$	0.57 %	50MV-500A
	1 m Ω	25 $\mu\Omega/\Omega$	CS-100
	10 m Ω	34 $\mu\Omega/\Omega$	Y5020
	100 m Ω	10 $\mu\Omega/\Omega$	VFP4
	1 Ω	0.33 $\mu\Omega/\Omega$	742A-1
	10 Ω	0.37 $\mu\Omega/\Omega$	742A-10
	25 Ω	0.93 $\mu\Omega/\Omega$	101-T
	100 Ω	0.42 $\mu\Omega/\Omega$	742A-100
	200 Ω	1.1 $\mu\Omega/\Omega$	VHA516-200
	400 Ω	0.77 $\mu\Omega/\Omega$	VHA516-400
	1 k Ω	0.51 $\mu\Omega/\Omega$	742A-1k
	10 k Ω	0.39 $\mu\Omega/\Omega$	742A-10k
	100 k Ω	1.3 $\mu\Omega/\Omega$	Ohm-Labs 105
	1 M Ω	2.0 $\mu\Omega/\Omega$	Ohm-Labs 106
	10 M Ω	3.0 $\mu\Omega/\Omega$	742A-10M
	100 M Ω	14 $\mu\Omega/\Omega$	9330-100M
	1 G Ω	150 $\mu\Omega/\Omega$	GT1 (1G)
	10 G Ω	220 $\mu\Omega/\Omega$	GT1 (10G)
	100 G Ω	0.16 %	GT1 (100G)
	1 T Ω	0.23 %	GT1 (1T)
	0 Ω	39 $\mu\Omega$	5720A or 5730A
	1 Ω	74 $\mu\Omega/\Omega$	
	1.9 Ω	74 $\mu\Omega/\Omega$	
	10 Ω	19 $\mu\Omega/\Omega$	
	19 Ω	19 $\mu\Omega/\Omega$	
	100 Ω	8.5 $\mu\Omega/\Omega$	
	190 Ω	8.5 $\mu\Omega/\Omega$	
	1 k Ω	7.0 $\mu\Omega/\Omega$	
	1.9 k Ω	7.0 $\mu\Omega/\Omega$	
	10 k Ω	7.0 $\mu\Omega/\Omega$	
	19 k Ω	7.0 $\mu\Omega/\Omega$	
	100 k Ω	8.5 $\mu\Omega/\Omega$	
	190 k Ω	8.5 $\mu\Omega/\Omega$	
	1 M Ω	14 $\mu\Omega/\Omega$	
	1.9 M Ω	15 $\mu\Omega/\Omega$	
	10 M Ω	29 $\mu\Omega/\Omega$	
	19 M Ω	36 $\mu\Omega/\Omega$	
	100 M Ω	85 $\mu\Omega/\Omega$	

Parameter/Equipment	Range	CMC ^{2, 5} (\pm)	Comments
Resistance ³ – Generate	(0 to 10.9999) Ω (11 to 32.9999) Ω (33 to 109.9999) Ω 110 Ω to 1.099 999 k Ω (1.1 to 10.999 99) k Ω (11 to 109.9999) k Ω 110 k Ω to 1.099 999 M Ω (1.1 to 3.299 999) M Ω (3.3 to 10.999 99) M Ω (11 to 32.999 99) M Ω (33 to 109.9999) M Ω (110 to 329.9999) M Ω (330 to 1100) M Ω	30 $\mu\Omega/\Omega + 0.76 \text{ m}\Omega$ 23 $\mu\Omega/\Omega + 1.1 \text{ m}\Omega$ 21 $\mu\Omega/\Omega + 1.1 \text{ m}\Omega$ 21 $\mu\Omega/\Omega + 1.5 \text{ m}\Omega$ 21 $\mu\Omega/\Omega + 15 \text{ m}\Omega$ 21 $\mu\Omega/\Omega + 0.15 \Omega$ 24 $\mu\Omega/\Omega + 1.5 \Omega$ 46 $\mu\Omega/\Omega + 23 \Omega$ 99 $\mu\Omega/\Omega + 38 \Omega$ 190 $\mu\Omega/\Omega + 1.9 \text{ k}\Omega$ 380 $\mu\Omega/\Omega + 2.3 \text{ k}\Omega$ 0.23 % + 76 k Ω 1.1 % + 380 k Ω	552X series calibrator

Parameter/Range	Frequency	CMC ^{2, 5} (\pm)	Comments
AC Voltage – Measure/Generate			
(0.6 to 2.2) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	1300 $\mu\text{V}/\text{V} + 0.99 \mu\text{V}$ 560 $\mu\text{V}/\text{V} + 0.99 \mu\text{V}$ 320 $\mu\text{V}/\text{V} + 0.99 \mu\text{V}$ 620 $\mu\text{V}/\text{V} + 1.5 \mu\text{V}$ 910 $\mu\text{V}/\text{V} + 1.9 \mu\text{V}$ 1800 $\mu\text{V}/\text{V} + 3.0 \mu\text{V}$ 2000 $\mu\text{V}/\text{V} + 6.1 \mu\text{V}$ 3800 $\mu\text{V}/\text{V} + 6.1 \mu\text{V}$	5790B with 5720A or 5730A
(1.9 to 7) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	650 $\mu\text{V}/\text{V} + 0.99 \mu\text{V}$ 280 $\mu\text{V}/\text{V} + 0.99 \mu\text{V}$ 160 $\mu\text{V}/\text{V} + 0.99 \mu\text{V}$ 310 $\mu\text{V}/\text{V} + 1.5 \mu\text{V}$ 460 $\mu\text{V}/\text{V} + 1.9 \mu\text{V}$ 910 $\mu\text{V}/\text{V} + 3.0 \mu\text{V}$ 1100 $\mu\text{V}/\text{V} + 6.1 \mu\text{V}$ 2700 $\mu\text{V}/\text{V} + 6.1 \mu\text{V}$	

Parameter/Range	Frequency	CMC ^{2, 5} (\pm)	Comments
AC Voltage – Measure/Generate (cont)			
(6 to 22) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	220 μ V/V + 0.99 μ V 140 μ V/V + 0.99 μ V 84 μ V/V + 0.99 μ V 160 μ V/V + 1.5 μ V 240 μ V/V + 1.9 μ V 620 μ V/V + 3.0 μ V 760 μ V/V + 6.1 μ V 2000 μ V/V + 6.1 μ V	5790B with 5720A or 5730A
(19 to 70) mV	(9.5 to 10) Hz (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	760 μ V/V + 1.1 μ V 180 μ V/V + 1.1 μ V 99 μ V/V + 1.1 μ V 53 μ V/V + 1.1 μ V 99 μ V/V + 1.5 μ V 200 μ V/V + 1.9 μ V 400 μ V/V + 3.0 μ V 520 μ V/V + 6.1 μ V 990 μ V/V + 6.1 μ V	
(60 to 220) mV	(9.5 to 10) Hz (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	760 μ V/V + 1.1 μ V 160 μ V/V + 1.1 μ V 66 μ V/V + 1.1 μ V 33 μ V/V + 1.1 μ V 56 μ V/V + 1.5 μ V 120 μ V/V + 1.9 μ V 210 μ V/V + 3.0 μ V 300 μ V/V + 6.1 μ V 910 μ V/V + 6.1 μ V	
(190 to 700) mV	(9.5 to 10) Hz (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	760 μ V/V + 1.1 μ V 160 μ V/V + 1.1 μ V 59 μ V/V + 1.1 μ V 29 μ V/V + 1.1 μ V 43 μ V/V + 1.5 μ V 64 μ V/V + 1.9 μ V 160 μ V/V + 3.0 μ V 260 μ V/V + 6.1 μ V 910 μ V/V + 6.1 μ V	

Parameter/Range	Frequency	CMC ^{2, 5} (\pm)	Comments
AC Voltage – Measure/Generate (cont)			
(0.6 to 2.2) V	(9.5 to 10) Hz (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	760 μ V/V 150 μ V/V 53 μ V/V 22 μ V/V 40 μ V/V 58 μ V/V 150 μ V/V 240 μ V/V 910 μ V/V	5790B with 5720A or 5730A
(1.9 to 7) V	(9.5 to 10) Hz (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	760 μ V/V 150 μ V/V 53 μ V/V 22 μ V/V 40 μ V/V 67 μ V/V 170 μ V/V 360 μ V/V 1100 μ V/V	
(6 to 22) V	(9.5 to 10) Hz (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	760 μ V/V 150 μ V/V 53 μ V/V 24 μ V/V 40 μ V/V 65 μ V/V 170 μ V/V 360 μ V/V 1100 μ V/V	
(19 to 70) V	(9.5 to 10) Hz (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	760 μ V/V 150 μ V/V 55 μ V/V 30 μ V/V 48 μ V/V 84 μ V/V 170 μ V/V 390 μ V/V 1100 μ V/V	

Parameter/Range	Frequency	CMC ^{2, 5} (±)	Comments
AC Voltage – Measure/Generate (cont)			
(60 to 220) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz	150 µV/V 55 µV/V 29 µV/V 59 µV/V 84 µV/V 200 µV/V 530 µV/V	5790B with 5720A or 5730A
(190 to 700) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz	150 µV/V 84 µV/V 36 µV/V 110 µV/V 650 µV/V	
(600 to 1050) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz	150 µV/V 84 µV/V 33 µV/V 110 µV/V 650 µV/V	
AC Voltage – Measure			
(1 to 12.12) mV	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz	290 µV/V + 1.1 µV 370 µV/V + 1.1 µV 380 µV/V + 1.1 µV 0.3 % + 1.1 µV 1 % + 4 µV 2 % + 4 µV	8588A
(1 to 121.2) mV	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz	88 µV/V + 0.5 µV 130 µV/V + 0.5 µV 230 µV/V + 1 µV 530 µV/V + 5 µV 0.21 % + 30 µV 1.1 % + 0.1 mV 1.6 % + 0.5 mV 4.1 % + 1 mV 8.4 % + 1 mV 16 % + 1 mV	

Parameter/Range	Frequency	CMC ^{2, 5} (\pm)	Comments
AC Voltage – Measure (cont)			
(0.01 to 1.212) V	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz	76 μ V/V + 5 μ V 130 μ V/V + 5 μ V 230 μ V/V + 10 μ V 530 μ V/V + 50 μ V 0.21 % + 0.3 mV 1 % + 1 mV 1.5 % + 5 mV 4 % + 10 mV 8.2 % + 10 mV 16 % + 10 mV	8588A
(0.1 to 12.12) V	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz	76 μ V/V + 50 μ V 130 μ V/V + 50 μ V 230 μ V/V + 100 μ V 530 μ V/V + 500 μ V 0.21 % + 3 mV 1 % + 10 mV 1.5 % + 50 mV 4 % + 100 mV 8.2 % + 100 mV 16 % + 100 mV	
(1 to 121.2) V	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz	90 μ V/V + 0.5 mV 110 μ V/V + 0.5 mV 230 μ V/V + 1 mV 590 μ V/V + 5 mV 0.37 % + 50 mV 1.1 % + 0.5 V	
(10 to 1050) V	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz	110 μ V/V + 25 mV 110 μ V/V + 25 mV 230 μ V/V + 25 mV 590 μ V/V + 100 mV	

Parameter/Range	Frequency	CMC ^{2, 5} (±)	Comments
AC Voltage ³ – Measure			
(1 to 12) mV	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz	0.034 % + 3.4 µV 0.023 % + 1.2 µV 0.034 % + 1.2 µV 0.11 % + 1.2 µV 0.57 % + 1.2 µV 4.5 % + 2.3 µV	3458A
(10 to 120) mV	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz	0.0079 % + 4.5 µV 0.0079 % + 2.3 µV 0.016 % + 2.3 µV 0.034 % + 2.3 µV 0.091 % + 2.3 µV 0.34 % + 11 µV 1.1 % + 11 µV 1.7 % + 11 µV	
(0.10 to 1.2) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz	0.0079 % + 45 µV 0.0079 % + 23 µV 0.016 % + 23 µV 0.034 % + 23 µV 0.091 % + 23 µV 0.34 % + 110 µV 1.1 % + 110 µV 1.7 % + 110 µV	
(1 to 12) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz	0.0079 % + 0.45 mV 0.0079 % + 0.23 mV 0.016 % + 0.23 mV 0.034 % + 0.23 mV 0.091 % + 0.23 mV 0.34 % + 1.1 mV 1.1 % + 1.1 mV 1.7 % + 1.1 mV	
(10 to 120) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz	0.023 % + 4.5 mV 0.023 % + 2.3 mV 0.023 % + 2.3 mV 0.04 % + 2.3 mV 0.14 % + 2.3 mV 0.45 % + 11 mV 1.7 % + 11 mV	

Parameter/Range	Frequency	CMC ^{2, 5} (\pm)	Comments
AC Voltage ³ – Measure (cont)			
(100 to 700) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.045 % + 45 mV 0.045 % + 23 mV 0.068 % + 23 mV 0.14 % + 23 mV 0.34 % + 23 mV	3458A
(0 to 10) kV (10 to 100) kV	60 Hz 60 Hz	0.14 % + 0.11 V 0.57 % + 1.1 V	Vitrek 4700 w/ HLV-150 probe
AC Voltage – Generate			
(0.2 to 2.2) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	210 μ V/V + 3.9 μ V 81 μ V/V + 3.9 μ V 70 μ V/V + 3.9 μ V 180 μ V/V + 3.9 μ V 420 μ V/V + 4.7 μ V 930 μ V/V + 9.3 μ V 1200 μ V/V + 19 μ V 2400 μ V/V + 19 μ V	5720A or 5730A w/ 5725A
(2.2 to 22) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	210 μ V/V + 3.9 μ V 81 μ V/V + 3.9 μ V 70 μ V/V + 3.9 μ V 180 μ V/V + 3.9 μ V 420 μ V/V + 4.7 μ V 930 μ V/V + 9.3 μ V 1200 μ V/V + 19 μ V 2400 μ V/V + 19 μ V	
(22 to 220) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	210 μ V/V + 12 μ V 81 μ V/V + 6.2 μ V 70 μ V/V + 6.2 μ V 180 μ V/V + 6.2 μ V 420 μ V/V + 16 μ V 700 μ V/V + 19 μ V 1200 μ V/V + 23 μ V 2200 μ V/V + 47 μ V	

Parameter/Range	Frequency	CMC ^{2, 5} (\pm)	Comments
AC Voltage – Generate (cont)			
(0.22 to 2.2) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	210 μ V/V + 39 μ V 78 μ V/V + 16 μ V 36 μ V/V + 7.8 μ V 66 μ V/V + 9.3 μ V 97 μ V/V + 31 μ V 330 μ V/V + 78 μ V 850 μ V/V + 190 μ V 1400 μ V/V + 470 μ V	5720A or 5730A w/ 5725A
(2.2 to 22) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	210 μ V/V + 390 μ V 78 μ V/V + 160 μ V 36 μ V/V + 54 μ V 66 μ V/V + 93 μ V 89 μ V/V + 190 μ V 240 μ V/V + 620 μ V 850 μ V/V + 1900 μ V 1200 μ V/V + 3100 μ V	
(22 to 220) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	210 μ V/V + 3.9 mV 78 μ V/V + 1.6 mV 47 μ V/V + 0.54 mV 74 μ V/V + 0.93 mV 130 μ V/V + 2.3 mV 780 μ V/V + 16 mV 4000 μ V/V + 39 mV 7000 μ V/V + 78 mV	
(220 to 1100) V	(15 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 30) kHz (30 to 50) kHz (50 to 100) kHz	62 μ V/V + 3.1 mV 62 μ V/V + 3.1 mV 97 μ V/V + 4.7 mV 280 μ V/V + 8.5 mV 280 μ V/V + 8.5 mV 1000 μ V/V + 35 mV	

Parameter/Range	Frequency	CMC ^{2, 5} (\pm)	Comments
AC Voltage ³ – Generate			
(1 to 32.999) mV	(10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	610 μ V/V + 4.6 μ V 110 μ V/V + 4.6 μ V 150 μ V/V + 4.6 μ V 760 μ V/V + 4.6 μ V 2700 μ V/V + 9.1 μ V 6100 μ V/V + 38 μ V	552X series calibrator
(33 to 329.999) mV	(10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	230 μ V/V + 6.1 μ V 110 μ V/V + 6.1 μ V 120 μ V/V + 6.1 μ V 270 μ V/V + 6.1 μ V 610 μ V/V + 24 μ V 1500 μ V/V + 53 μ V	
(0.33 to 3.299 99) V	(10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	230 μ V/V + 38 μ V 110 μ V/V + 46 μ V 140 μ V/V + 46 μ V 230 μ V/V + 38 μ V 530 μ V/V + 95 μ V 1800 μ V/V + 460 μ V	
(3.3 to 32.9999) V	(10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz	230 μ V/V + 490 μ V 110 μ V/V + 460 μ V 180 μ V/V + 460 μ V 270 μ V/V + 460 μ V 680 μ V/V + 1200 μ V	
(33 to 329.999) V	45 Hz to 1 kHz (1 to 10) kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz	140 μ V/V + 1500 μ V 150 μ V/V + 4600 μ V 190 μ V/V + 4600 μ V 230 μ V/V + 4600 μ V 1500 μ V/V + 38 000 μ V	
(330 to 1020) V	45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	230 μ V/V + 7600 μ V 190 μ V/V + 7600 μ V 230 μ V/V + 7600 μ V	

Parameter/Range	Frequency	CMC ^{2, 5, 6} (\pm)	Comments
AC Voltage – Generate Wideband AC Voltage Flatness:			
1.1 mV	(10 to 30) Hz 30 Hz to 120 kHz 120 kHz to 2 MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz	0.23 % 0.078 % 0.16 % + 2.3 μ V 0.31 % + 2.3 μ V 0.47 % + 2.3 μ V 1.2 % + 12 μ V	5720A wideband
3 mV	(10 to 30) Hz 30 Hz to 120 kHz 120 kHz to 2 MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz	0.23 % 0.078 % 0.078 % + 2.3 μ V 0.23 % + 2.3 μ V 0.39 % + 2.3 μ V 1.2 % + 2.3 μ V	
11 mV to 3.5 V	(10 to 30) Hz 30 Hz to 120 kHz 120 kHz to 2 MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz	0.23 % 0.078 % 0.078 % + 2.3 μ V 0.16 % + 2.3 μ V 0.31 % + 2.3 μ V 0.78 % + 2.3 μ V	
AC Voltage Flatness Relative to 1 kHz – Measure/Generate			
(0.6 to 2.2) mV	(10 to 30) Hz 30 Hz to 120 kHz 120 kHz to 2 MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz (30 to 50) MHz	0.11 % 0.057 % 0.080 % + 1.1 μ V 0.19 % + 1.1 μ V 0.34 % + 1.1 μ V 0.80 % + 2.3 μ V 1.1 % + 2.3 μ V	5790B & 5720A wideband

Parameter/Range	Frequency	CMC ^{2, 5, 6} (±)	Comments
AC Voltage Flatness Relative to 1 kHz – Measure/Generate (cont)			
(1.9 to 7) mV	(10 to 30) Hz 30 Hz to 120 kHz 120 kHz to 2 MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz (30 to 50) MHz	0.11 % 0.057 % 0.080 % + 1.1 µV 0.11 % + 1.1 µV 0.19 % + 1.1 µV 0.42 % + 1.1 µV 0.57 % + 1.1 µV	5790B & 5720A wideband
(6 to 22) mV	(10 to 30) Hz 30 Hz to 120 kHz 120 kHz to 2 MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz (30 to 50) MHz	0.11 % 0.057 % 0.080 % 0.11 % 0.19 % 0.42 % 0.68 %	
(19 to 70) mV	(10 to 30) Hz 30 Hz to 2 MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz (30 to 50) MHz	0.11 % 0.057 % 0.11 % 0.17 % 0.40 % 0.68 %	
(60 to 220) mV	(10 to 30) Hz 30 Hz to 500 kHz 500 kHz to 2 MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz (30 to 50) MHz	0.11 % 0.046 % 0.057 % 0.11 % 0.17 % 0.40 % 0.68 %	
190 mV to 7 V	(10 to 30) Hz 30 Hz to 500 kHz 500 kHz to 2 MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz (30 to 50) MHz	0.11 % 0.034 % 0.057 % 0.11 % 0.17 % 0.40 % 0.68 %	

Parameter/Range	Frequency	CMC ^{2, 5, 6} (±)	Comments
AC Voltage Flatness Relative to 50 kHz – Measure/Generate			
5 mV _{p-p} to 5.5 V _{p-p}	9 kHz to 1.2 GHz (> 1.2 to 18) GHz	0.62 % 1.7 %	E9304A or E9304A-H18 w/ signal generator
AC Current – Measure/Generate			
(10 to 70) µA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 30) kHz	320 µA/A 250 µA/A 190 µA/A 210 µA/A	5790A + AC shunts & 5720A or 5730A w/ 5725A or 552X series calibrator
70 µA to < 0.33 A	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 30) kHz	190 µA/A 120 µA/A 100 µA/A 110 µA/A	
(0.33 to 20) A	(10 to 20) Hz (20 to 40) Hz 40 Hz to 10 kHz	190 µA/A 120 µA/A 110 µA/A	
AC Current – Measure			
(0.1 to 20) A	(40 to 310) Hz 310 Hz to 1 kHz (1 to 3) kHz (3 to 5) kHz	240 µA/A + 42 µA 330 µA/A + 40 µA 600 µA/A + 35 µA 870 µA/A + 26 µA	Y5020, 8588A
(0.202 to 20.2) µA	1 Hz to 2 kHz (2 to 10) kHz	0.21 % + 2.5 nA 0.21 % + 2.5 nA	8588A
(20.2 to 202) µA	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz	280 µA/A + 5 nA 530 µA/A + 5 nA 740 µA/A + 5 nA	
(0.202 to 2.02) mA	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz	280 µA/A + 50 nA 530 µA/A + 50 nA 740 µA/A + 50 nA	
(2.02 to 20.2) mA	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz	280 µA/A + 0.5 µA 530 µA/A + 0.5 µA 740 µA/A + 0.5 µA	

Parameter/Range	Frequency	CMC ^{2, 5} (±)	Comments
AC Current – Measure (cont)			
(20.2 to 202) mA	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz	280 µA/A + 5 µA 520 µA/A + 5 µA 740 µA/A + 5 µA	8588A
(0.202 to 2.02) A	1 Hz to 2 kHz (2 to 10) kHz	300 µA/A + 0.1 mA 550 µA/A + 0.1 mA	
(2.02 to 20.2) A	10 Hz to 2 kHz (2 to 10) kHz	840 µA/A + 0.5 mA 840 µA/A + 0.5 mA	
(20.2 to 30.2) A	10 Hz to 2 kHz (2 to 10) kHz	840 µA/A + 12 mA 0.13 % + 12 mA	
(1 to 100) A	60 Hz	110 µA/A + 0.5 mA	CS-100, 8588A
AC Current ³ – Measure			
(5 to 100) µA	(10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz	0.45 % + 0.034 µA 0.17 % + 0.034 µA 0.068 % + 0.034 µA	3458A
(0.05 to 1) mA	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	0.45 % + 0.23 µA 0.17 % + 0.23 µA 0.068 % + 0.23 µA 0.034 % + 0.23 µA	
(0.5 to 10) mA	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	0.45 % + 2.3 µA 0.17 % + 2.3 µA 0.068 % + 2.3 µA 0.034 % + 2.3 µA	
(5 to 100) mA	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	0.45 % + 23 µA 0.17 % + 23 µA 0.068 % + 23 µA 0.034 % + 23 µA	
(0.05 to 1) A	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	0.45 % + 230 µA 0.18 % + 230 µA 0.091 % + 230 µA 0.11 % + 230 µA	

Parameter/Range	Frequency	CMC ^{2, 5} (±)	Comments
AC Current – Generate			
220 µA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	220 µA/A + 16 nA 140 µA/A + 9.3 nA 100 µA/A + 7.8 nA 250 µA/A + 12 nA 850 µA/A + 62 nA	5720A or 5730A w/ 5725A
2.2 mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	220 µA/A + 39 nA 140 µA/A + 31 nA 100 µA/A + 31 nA 170 µA/A + 100 nA 850 µA/A + 620 nA	
22 mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	220 µA/A + 390 nA 140 µA/A + 310 nA 100 µA/A + 310 nA 170 µA/A + 540 nA 850 µA/A + 4700 nA	
220 mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	220 µA/A + 3.9 µA 140 µA/A + 3.1 µA 100 µA/A + 2.3 µA 170 µA/A + 3.1 µA 850 µA/A + 9.3 µA	
2.2 A	20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	230 µA/A + 31 µA 360 µA/A + 78 µA 5400 µA/A + 160 µA	
11 A	40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	310 µA/A + 130 µA 660 µA/A + 290 µA 2600 µA/A + 580 µA	
AC Current ³ – Generate			
(29 to 329.99) µA	(10 to 20) Hz (20 to 45) Hz 45 to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.15 % + 0.076 µA 0.11 % + 0.076 µA 0.095 % + 0.076 µA 0.23 % + 0.11 µA 0.61 % + 0.15 µA 1.2 % + 0.3 µA	552X series calibrator

Parameter/Range	Frequency	CMC ^{2, 5} (±)	Comments
AC Current ³ – Generate (cont)			
(0.33 to 3.2999) mA	(10 to 20) Hz (20 to 45) Hz 45 to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.15 % + 0.11 µA 0.095 % + 0.11 µA 0.076 % + 0.11 µA 0.15 % + 0.15 µA 0.38 % + 0.23 µA 0.76 % + 0.46 µA	552X series calibrator
(3.3 to 32.999) mA	(10 to 20) Hz (20 to 45) Hz 45 to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.14 % + 1.5 µA 0.068 % + 1.5 µA 0.03 % + 1.5 µA 0.061 % + 1.5 µA 0.15 % + 2.3 µA 0.3 % + 3 µA	
(33 to 329.99) mA	(10 to 20) Hz (20 to 45) Hz 45 to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.14 % + 15 µA 0.068 % + 15 µA 0.03 % + 15 µA 0.076 % + 38 µA 0.15 % + 76 µA 0.3 % + 150 µA	
(0.33 to 1.099 99) A	(10 to 45) Hz 45 to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.14 % + 76 µA 0.038 % + 76 µA 0.46 % + 760 µA 1.9 % + 3800 µA	
(1.1 to 2.999 99) A	(10 to 45) Hz 45 to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.14 % + 76 µA 0.046 % + 76 µA 0.46 % + 760 µA 1.9 % + 3800 µA	
(3 to 10.9999) A	(45 to 100) Hz 100 to 1 kHz (1 to 5) kHz	0.046 % + 1500 µA 0.076 % + 1500 µA 2.3 % + 1500 µA	
(11 to 20.5) A	(45 to 100) Hz 100 to 1 kHz (1 to 5) kHz	0.091 % + 3800 µA 0.11 % + 3800 µA 2.3 % + 3800 µA	

Parameter/Range	Frequency	CMC ^{2, 5, 6} (\pm)	Comments
AC Current ³ – Generate (cont)			
Clamp-On Meters:			
Toroidal Type: (20 to 54.999) A (55 to 149.999) A (150 to 1025) A	(45 to 65) Hz	0.31 % + 0.054 A 0.34 % + 0.042 A 0.34 % + 0.13 A	552X series w/ 5500A/coil
(20 to 54.999) A (55 to 149.999) A (150 to 400) A	(65 to 440) Hz	0.93 % + 0.054 A 0.94 % + 0.046 A 1.2 % + 0.22 A	
Non-Toroidal Type: (20 to 149.999) A (150 to 549.999) A (550 to 1025) A	(45 to 65) Hz	0.65 % + 0.30 A 0.66 % + 1.1 A 0.65 % + 1.1 A	
(20 to 149.999) A (150 to 400) A	(65 to 440) Hz	1.2 % + 0.30 A 1.4 % + 1.1 A	
Capacitance – Measure ¹²			
1 pF 10 pF 100 pF 400 pF to 25 μ F 100 μ F 1000 μ F	12 Hz to 100 kHz	4.6 % 0.47 % 0.057 % 0.024 % 0.057 % 0.47 %	Genrad 1689 CMC is stated at 1kHz ¹³

Parameter/Equipment	Range	CMC ^{2, 5} (\pm)	Comments
Capacitance – Measure/Generate	0 F (0 to 2.02) nF (1.8 to 20.2) nF (18 to 202) nF (0.18 to 2.02) μ F (1.8 to 20.2) μ F	0.23 pF 0.19 % + 1 pF 810 μ F/F + 2 pF 490 μ F/F + 10 pF 420 μ F/F + 0.1 nF 420 μ F/F + 1 nF	Open input 8588A w/ 552X series calibrator

Parameter/Equipment	Range	CMC ^{2, 5, 6} (\pm)	Comments
Capacitance – Measure/Generate (cont)	(18 to 202) μ F (0.18 to 2.02) mF (1.8 to 20.2) mF (18 to 202) mF	620 μ F/F + 10 nF 620 μ F/F + 0.1 μ F 720 μ F/F + 1 μ F 720 μ F/F + 10 μ F	8588A w/ 552X series calibrator
Capacitance – Measure DC Charge Method	200 μ F to 110 mF	0.020 %	3458A w/ 5720A or 5730A

Parameter/Range	Frequency	CMC ^{2, 5, 6} (\pm)	Comments
Capacitance ³ – Generate			
(0.22 to 1.0999) nF (1.1 to 3.2999) nF (3.3 to 10.999) nF (11 to 109.999) nF (110 to 329.999) nF (0.33 to 1.099 99) μ F (1.1 to 3.299 99) μ F (3.3 to 10.9999) μ F (11 to 32.9999) μ F (33 to 109.999) μ F (110 to 329.999) μ F (0.33 to 1.099 99) mF (1.1 to 3.2999) mF (3.3 to 10.9999) mF (11 to 32.9999) mF (33 to 110) mF	10 Hz to 10 kHz 10 Hz to 3 kHz 10 Hz to 1 kHz 10 Hz to 1 kHz 10 Hz to 1 kHz (10 to 600) Hz (10 to 300) Hz (10 to 150) Hz (10 to 120) Hz (10 to 80) Hz (0 to 50) Hz (0 to 20) Hz (0 to 6) Hz (0 to 2) Hz (0 to 0.6) Hz (0 to 0.2) Hz	0.38 % + 7.6 pF 0.38 % + 7.6 pF 0.19 % + 7.6 pF 0.19 % + 76 pF 0.19 % + 230 pF 0.19 % + 0.76 nF 0.19 % + 2.3 nF 0.19 % + 7.6 nF 0.3 % + 23 nF 0.34 % + 76 nF 0.34 % + 230 nF 0.34 % + 0.76 μ F 0.34 % + 2.3 μ F 0.34 % + 7.6 μ F 0.57 % + 23 μ F 0.84 % + 76 μ F	552X series calibrator
Inductance – Measure ¹²			
100 μ H 1 mH to 10 H	100 Hz to 10 kHz	0.13 % 0.024 %	Genrad 1689 CMC is stated at 1kHz ¹³

Parameter/Range	Frequency	CMC ^{2, 5, 6} (±)	Comments
AC Resistance – Measure ¹² 0.1 Ω 1 Ω 6.25 Ω to 410 kΩ 1 MΩ 10 MΩ 100 MΩ	12 Hz to 100 kHz	0.72 % 0.083 % 0.024 % 0.040 % 0.29 % 2.8 %	Genrad 1689 CMC is stated at 1kHz ¹³

Parameter/Equipment	Range	CMC ² (±)	Comments
Electrical Simulation of Thermocouples & Thermocouple Indicating Devices ³ – Type B (600 to 800) °C (800 to 1000) °C (1000 to 1550) °C (1550 to 1820) °C Type E (-250 to -100) °C (-100 to -25) °C (-25 to 350) °C (350 to 650) °C (650 to 1000) °C Type J (-210 to -100) °C (-100 to -30) °C (-30 to 150) °C (150 to 760) °C (760 to 1200) °C Type K (-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 1000) °C (1000 to 1372) °C Type N (-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 410) °C (410 to 1300) °C			552X series calibrator

Parameter/Equipment	Range	CMC ^{2, 5, 6} (±)	Comments
Electrical Simulation of Thermocouples & Thermocouple Indicating Devices ³ – (cont)			
Type R	(0 to 250) °C (250 to 400) °C (400 to 1000) °C (1000 to 1767) °C	0.43 °C 0.27 °C 0.25 °C 0.30 °C	552X series calibrator
Type S	(0 to 250) °C (250 to 1000) °C (1000 to 1400) °C (1400 to 1767) °C	0.36 °C 0.27 °C 0.28 °C 0.35 °C	
Type T	(-250 to -150) °C (-150 to 0) °C (0 to 120) °C (120 to 400) °C	0.48 °C 0.18 °C 0.12 °C 0.11 °C	
Electrical Simulation of RTD Indicators & Indicating Systems ³ –			
Pt 385, 100 Ω	(-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C (400 to 630) °C (630 to 800) °C	0.038 °C 0.038 °C 0.053 °C 0.068 °C 0.076 °C 0.091 °C 0.18 °C	552X series calibrator
Oscilloscopes ³ –			
Amplitude – DC Signal: 50 Ω Load	(0 to 6.6) V	0.19 % + 30 µV	552X/SC1100
1 MΩ Load	(0 to 130) V	0.038 % + 30 µV	
Amplitude – Square Wave: 50 Ω Load	1 mV to 6.6 V _{p-p} 10 Hz to 10 kHz	0.19 % + 30 µV	
1 MΩ Load	1 mV to 130 V _{p-p} 10 Hz to 10 kHz	0.076 % + 30 µV	

Parameter/Equipment	Range	CMC ^{2, 5, 6} (\pm)	Comments
Oscilloscopes ³ – (cont)			
Bandwidth Flatness	5 mV to 5.5 V _{p-p} 50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz	1.1 % + 76 μ V 1.5 % + 76 μ V 3.0 % + 76 μ V	552X/SC1100
	5 mV to 3.5 V _{p-p} (600 to 1100) MHz	3.8 % + 76 μ V	
Time Marker	5 s to 50 ms 20 ms to 1 ns	19 μ s/s + 760 μ s/s * t 1.9 μ s/s	t = time in seconds
Resistance – Measure	(40 to 60) Ω (0.5 to 1.5) M Ω	0.076 % 0.076 %	
Phase Angle – Measure			
(0 to 360) $^{\circ}$	5 Hz to 2 kHz (2 to 5) kHz (5 to 10) kHz (10 to 50) kHz (50 to 100) kHz (100 to 200) kHz	0.023 $^{\circ}$ 0.034 $^{\circ}$ 0.046 $^{\circ}$ 0.057 $^{\circ}$ 0.11 $^{\circ}$ 0.23 $^{\circ}$	Clarke-Hess 6000A phase meter
Phase Angle – Generate			
(0 to 360) $^{\circ}$	1 Hz to 1 kHz (1 to 6.25) kHz (6.25 to 50) kHz (50 to 200) kHz	0.0057 $^{\circ}$ 0.011 $^{\circ}$ 0.017 $^{\circ}$ 0.046 $^{\circ}$	Clarke-Hess 5600 phase standard
Rise Time – Measure	25 mV _{p-p} to 2.5 V _{p-p}	10 ps	Tektronix TDS820

III. Electrical – RF/Microwave

Parameter/Range	Frequency	CMC ^{2, 5} (\pm)	Comments
RF Power – Generate (+23.9 to -56) dBm	100 kHz to 10 MHz (> 10 to 60) MHz (> 60 to 80) MHz	0.23 dB 0.43 dB 0.57 dB	Keysight 33611A
RF Power – Measure Flatness: (-35 to +20) dBm (-60 to +20) dBm Absolute: (-35 to +20) dBm	1 kHz to 100 MHz 9 kHz to 1.2 GHz (> 1.2 to 6) GHz (6 to 18) GHz DC to 100 MHz 100 MHz to 2.4 GHz (2.4 to 8) GHz (8 to 12.4) GHz (12.4 to 18) GHz	0.020 dB 0.054 dB 0.15 dB 0.15 dB 0.048 dB 0.059 dB 0.068 dB 0.080 dB 0.11 dB	NRP18T + 10dB attenuator E9304A / E9304A-H18 E9304A-H18 NRP18T

IV. Mechanical

Parameter/Equipment	Range	CMC ^{2, 4} (\pm)	Comments
Mass	0.0009 to 0.05 kg (> 0.05 to 0.2) kg (> 0.2 to 2) kg (> 2 to 30) kg	(0.09 + 1.1M) mg (0.77 + 0.3M) mg (0.77 + 0.72M) mg (12 + 0.4M) mg	ASTM class 1 & 2 weights, mass comparators
Dynamometer	(0.5 to 2000) g	(0.11 + 0.0044M) g	ASTM Class 5 weights

Parameter/Equipment	Range	CMC ^{2, 4, 6} (\pm)	Comments
Torque Testers ¹⁰	(0.000 33 to 7.5) lbf·ft	(0.24 – 0.024 T) %	Torque arm, ASTM Class 5 & 6 weights
Scales & Balances ^{3, 9}	5 mg to 0.0005 kg (> 0.0005 to 10) kg (> 10 to 30) kg 20 mg to 0.1 kg (> 0.1 to 2) kg (> 2 to 29) kg (> 29 to 230) kg	0.014 mg (0.034 + 6.9 M) mg (29 + 4 M) mg (1 + 150 M – 410 M^2) mg (16 + 47 M – 4.0 M^2) mg (27 + 26 M – 0.42 M^2) mg (2000 + 45 M – 0.07 M^2) mg	ASTM Class 1 & 2 weights ASTM Class 4, 5 & 6 weights
Hydraulic & Pneumatic Gage Pressure ^{3, 9, 10}	(0 to 40) in H ₂ O (0 to 20) psig (20 to 10 000) psig	0.04 in H ₂ O 0.02 psi 0.10 %	477B-2 Crystal gauge
Hydraulic Gage Pressure ^{9, 10}	(60 to 200) psig (200 to 1500) psig (1500 to 15 000) psig	0.016 psi 0.008 psi + 0.0032 % 0.0028 %	Ruska 2485 DWT
Torque Tools ^{3, 9, 10}	(> 0.875 to 10) lbf·in (> 10 to 100) lbf·in (> 25 to 250) lbf·in (> 5 to 50) lbf·ft (> 25 to 250) lbf·ft (> 100 to 1000) lbf·ft	(5 – 5.3 T) % 0.6 % 0.65 % 0.6 % 0.53 % (0.56 – (0.03 · 10 ⁻³) T) %	HIOS HP-10 50617.LOG 50619.LOG 50836.LOG 50702.LOG 50773.LOG
Load Cells / Force Gauges ¹⁰			
Compression	Up to 1000 lbf (> 1000 to 60 000) lbf	0.43 lbf + 0.026 % 0.75 lbf + 0.053 %	Load cells
Tension	Up to 25 000 lbf (> 25 000 to 60 000) lbf	0.43 lbf + 0.055 % 9.5 lbf + 0.05 %	Load cells
Torque Testers ¹⁰	(0.000 33 to 7.5) lbf·ft	(0.24 – 0.024 T) %	Torque arm, ASTM Class 5 & 6 weights

Parameter/Equipment	Range	CMC ^{2, 4, 6} (\pm)	Comments
Pneumatic Absolute/Gage Pressure ^{9, 10}	(1.5 to 15) psia (-50 to 50) in·H ₂ O (0 to 7.5) psig (> 7.5 to 15) psig (-13.2 to 30) psig (> 30 to 100) psig (0 to 300) psig (> 300 to 1000) psig (0 to 900) psig (> 900 to 3000) psig (0 to 3000) psig (> 3000 to 10 000) psig	0.0012 psi 0.005 in·H ₂ O + 0.01 % 0.000 75 psi 0.01 % 0.003 psi 0.01 % 0.03 psi 0.01 % 0.09 psi 0.01 % 0.3 psi 0.01 %	Paroscientific model 745 PM500 BG14K PM500 G100K PM600 A700K PM600 A7M PM600 A20M PM600 A70M
Force Gauges ^{9, 10}	(0.002 to 400) lbf	0.024 %	ASTM Class 5 & 6 Weights
Aircraft Scales ⁹	(500 to 5000) lbf (> 5000 to 25 000) lbf (> 25 000 to 60 000) lbf	2.9 lbf + 0.029 % 1.8 lbf + 0.052 % 10 lbf + 0.052 %	Load cells
Tensiometers ^{9, 10}	Up to 50 lbf (> 50 to 900) lbf	1.1 lbf 0.76 lbf + 0.7 %	Class 6 weights, tensiometer calibrator

V. Thermodynamics

Parameter/Equipment	Range	CMC ^{2, 6} (\pm)	Comments
Thermocouple Probes (E, J, K, T)	23 °C (room temperature)	0.011 °C	5615, 3458A, 8588A, lag bath
IR Thermometers	(35 to 500) °C	0.25 °C + 0.21 %	Fluke 4181 Emissivity = 0.95 λ = (8 to 14) μm

Parameter/Equipment	Range	CMC ^{2, 6} (\pm)	Comments
Thermocouple Thermometers (E, J, K, T)	23 °C (room temperature)	0.032 °C	5615, 3458A, lag bath
Thermometers	(-50 to 20) °C (20 to 300) °C	0.19 °C $0.03 + (0.56 \cdot 10^{-4})T$ °C	Fluke 1502A - RTD T is temperature in °C

VI. Time & Frequency

Parameter/Equipment	Range	CMC ^{2, 8} (\pm)	Comments
Frequency – Measure	0.1 Hz to 3 GHz (1 to 40) Hz 40 Hz to 10 MHz	470 pHz/Hz 0.057 % 0.011 %	53131A w/ 910R & signal generator 3458A
Frequency ³ – Measuring Equipment	0.01 Hz to 2 MHz 50 kHz to 1100 MHz 1 Hz to 80 MHz	2 μHz/Hz + 3.9 μHz 2 μHz/Hz 1.2 μHz/Hz	Fluke 552X Fluke 552X/SC1100 Keysight 33611A
Frequency – Fixed Point	10 MHz	3.9 pHz/Hz	Fluke 910R
Stopwatches & Timers	(1 to 86 400) s	1.8 ms	33611A, 53131A & 910R

¹ This laboratory offers commercial calibration service and field calibration service where noted.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards of nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the Calibration and Measurement Capability Uncertainty (CMC) found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the uncertainty introduced by the item being calibrated, (e.g., resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

⁴ In the statement of CMC, L is the numerical value of the nominal length of the device measured in inches for imperial units, or in millimeters for metric units. M is the numerical value of the nominal weight measured in kg (Mass, Scales). T is the numerical value of the nominal torque measured in lbf·ft.

⁵ The stated measured values are determined using the indicated instrument (see Comments). This capability is suitable for the calibration of the devices intended to measure or generate the measured value in the ranges indicated. CMC's are expressed as either a specific value that covers the full range or as a percent or fraction of the reading plus a fixed floor specification.

⁶ In the statement of CMC, percentages are to be read as percent of reading, unless otherwise noted.

⁷ This scope meets A2LA's *P112 Flexible Scope Policy*.

⁸ The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.

⁹ The contributions attributed to the repeatability of the "best existing device" are not included in the CMC claim.

¹⁰ The contributions attributed to the resolution of the "best existing device" are not included in the CMC claim.

¹² Uncertainty components that can be reasonably attributed to the Unit Under Test have not been utilized in the calculation of the CMC value for this measurement parameter.

¹³ Measurement uncertainty at intermediate values is calculated using the Manufacturers Limits of Error Calculator



Accredited Laboratory

A2LA has accredited

PYLON ELECTRONICS INC.

Ottawa, Ontario, CANADA

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 General requirements for the competence of testing and calibration laboratories. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system
(refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 1st day of October 2024.

A handwritten signature in blue ink, appearing to read "Trace McInturff".

Mr. Trace McInturff, Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 3630.01
Valid to September 30, 2026

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.